

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.9207 µl	10 µl	0.07932 µl	± 0.2 µl	±	5.1E-02 µl