

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(M.P.E)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0033 ml	1 ml	-0.00329 ml	± 0.007 ml	±	8.9E-02 µl
4.9903 ml	5 ml	0.00969 ml	± 0.030 ml	±	2.6E-01 µl
10.0086 ml	10 ml	-0.00858 ml	± 0.05 ml	±	5.0E-01 µl