

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.982 °C	30.1 °C	0.118 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.732 %HR	32.2 %HR	2.468 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.026 %HR	51.1 %HR	2.074 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.161 %HR	68.1 %HR	-2.061 %HR	± 3.0 %HR	± 1.9E+00 %HR