

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
999 RPM	1000 RPM	1 RPM	± 2 RPM	±	8.2E-02 RPM
1998 RPM	2000 RPM	2 RPM	± 2 RPM	±	1.1E-01 RPM
3508 RPM	3500 RPM	-8 RPM	± 2 RPM	±	1.1E-01 RPM
4015 RPM	4000 RPM	-15 RPM	± 2 RPM	±	1.1E-01 RPM