

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.8 mmHg	50 mmHg	0.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg
99.9 mmHg	100 mmHg	0.1 mmHg	± 3 mmHg	±	1.2E+00 mmHg
150.1 mmHg	150 mmHg	-0.1 mmHg	± 3 mmHg	±	1.2E+00 mmHg
200.2 mmHg	200 mmHg	-0.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg
249.8 mmHg	250 mmHg	0.2 mmHg	± 3 mmHg	±	1.2E+00 mmHg
299.7 mmHg	300 mmHg	0.3 mmHg	± 3 mmHg	±	1.2E+00 mmHg