

| Tipo de Medida (Measurement Type) | Patrón (Standard) | UBP (UT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |
|-----------------------------------|-------------------|----------|---------------|-------------------|-----------------------------|
| Profundidad / (Depth)             | 1.00 mm           | 1.0 mm   | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 1.50 mm           | 1.5 mm   | 0 µm          | ± 30 µm           | ± 5.8E+00 µm                |
| Profundidad / (Depth)             | 3.00 mm           | 3.0 mm   | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 5.00 mm           | 5.0 mm   | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 10.00 mm          | 10.0 mm  | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 15.00 mm          | 15.0 mm  | 10 µm         | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 20.00 mm          | 20.0 mm  | 10 µm         | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 25.00 mm          | 25.0 mm  | 10 µm         | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 30.00 mm          | 30.0 mm  | 10 µm         | ± 30 µm           | ± 6.1E+00 µm                |