

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |              |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|--------------|
| 50.7 mmHg                | 50 mmHg          | -0.7 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |
| 100.4 mmHg               | 100 mmHg         | -0.4 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |
| 150.7 mmHg               | 150 mmHg         | -0.7 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |
| 200.8 mmHg               | 200 mmHg         | -0.8 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |
| 250.7 mmHg               | 250 mmHg         | -0.7 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |
| 301.1 mmHg               | 300 mmHg         | -1.1 mmHg            | ± 3 mmHg                    | ±                                  | 1.2E+00 mmHg |