

Certificado No. (Certificate #):56926Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-05-14Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.0000 mm	1.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
2.0000 mm	2.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
4.0000 mm	4.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
7.0000 mm	7.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
8.0000 mm	8.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
10.0000 mm	10.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
13.0000 mm	13.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
16.0000 mm	16.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
19.0000 mm	19.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
22.0000 mm	22.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
25.0000 mm	25.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm