

Certificado No. (Certificate #):56927Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-05-15Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25.0000 mm	25.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
28.0000 mm	28.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
30.0000 mm	30.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
32.0000 mm	32.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
34.0000 mm	34.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
37.5000 mm	37.5000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
39.0000 mm	39.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
43.0000 mm	43.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
46.0000 mm	46.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
48.0000 mm	48.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm
50.0000 mm	50.0000 mm	0.0000 µm	± 4 µm	±	5.8E+00 µm