

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
75.0000 mm	75.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
78.0000 mm	78.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
81.0000 mm	81.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
82.0000 mm	82.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
84.0000 mm	84.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
87.0000 mm	87.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
90.0000 mm	90.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
93.0000 mm	93.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
96.0000 mm	96.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
99.0000 mm	99.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm
100.0000 mm	100.0000 mm	0.0000 µm	± 5 µm	±	5.8E+00 µm