

Certificado No. (Certificate #):56929Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-05-14Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
125.0000 mm	125.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
128.0000 mm	128.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
131.0000 mm	131.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
134.0000 mm	134.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
137.0000 mm	137.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
140.0000 mm	140.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
142.0000 mm	142.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
143.0000 mm	143.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
146.0000 mm	146.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
149.0000 mm	149.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm
150.0000 mm	150.0000 mm	0.0000 µm	± 6 µm	±	5.8E+00 µm