

Certificado No. (Certificate #):		56930		Equipo (Instrument):		Micrómetro de Interior			
Fecha de Calibración (Calibration Date):		2025-05-14		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
50.0000 mm		49.999 mm		-1.0000 µm		± 7 µm		± 5.8E+00 µm	
55.0000 mm		55.000 mm		0.0000 µm		± 7 µm		± 5.8E+00 µm	
60.0000 mm		60.000 mm		0.0000 µm		± 7 µm		± 5.8E+00 µm	
75.0000 mm		75.000 mm		0.0000 µm		± 7 µm		± 5.8E+00 µm	
100.0000 mm		100.000 mm		0.0000 µm		± 7 µm		± 5.8E+00 µm	
110.0000 mm		110.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	
125.0000 mm		125.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	
150.0000 mm		150.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	
170.0000 mm		170.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	
180.0000 mm		180.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	
200.0000 mm		200.000 mm		0.0000 µm		± 10 µm		± 5.8E+00 µm	