

Certificado No. (Certificate #):56930Equipo (Instrument):Micrómetro de Interior

Fecha de Calibración (Calibration Date):2025-05-20Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.0000 mm	50.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
55.0000 mm	55.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
60.0000 mm	60.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
75.0000 mm	75.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
100.0000 mm	100.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
110.0000 mm	110.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
125.0000 mm	125.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
150.0000 mm	150.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
170.0000 mm	170.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
180.0000 mm	180.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
200.0000 mm	200.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm