

Certificado No. (Certificate #):56930Equipo (Instrument):Micrómetro de Interior

Fecha de Calibración (Calibration Date):2025-05-20Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.0000 mm	50.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
75.0000 mm	75.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
100.0000 mm	100.000 mm	0.0000 µm	± 7 µm	±	5.8E+00 µm
125.0000 mm	125.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
150.0000 mm	150.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
175.0000 mm	175.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
200.0000 mm	200.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
225.0000 mm	225.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
250.0000 mm	250.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
270.0000 mm	270.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm
280.0000 mm	280.000 mm	0.0000 µm	± 10 µm	±	5.8E+00 µm