

Certificado No. (Certificate #):	56932	Equipo (Instrument): Profundímetro			
Fecha de Calibración (Calibration Date):	2025-05-21	Marca (Brand): Fowler			
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	0.050 in	0.050 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.100 in	0.100 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.110 in	0.110 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.150 in	0.150 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.200 in	0.200 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.300 in	0.300 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.400 in	0.400 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	0.500 in	0.500 in	0.000 in	± 0.0019685 in	± 5.8E-04 in
Profundidad / (Depth)	1.000 in	1.000 in	0.000 in	± 0.0019685 in	± 5.8E-04 in