

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0802 µl	10 µl	-0.08023 µl	± 0.50 µl	±	5.4E-02 µl
50.3510 µl	50 µl	-0.35103 µl	± 0.50 µl	±	1.2E-01 µl
100.3913 µl	100 µl	-0.39128 µl	± 0.50 µl	±	1.4E-01 µl