

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.1206 µl	100 µl	-0.12061 µl	± 0.50 µl	±	#REF! #REF!
500.2522 µl	500 µl	-0.25215 µl	± 0.50 µl	±	#REF! #REF!
1,000.3038 µl	1000 µl	-0.30380 µl	± 0.50 µl	±	#REF! #REF!