

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.9248 µl	10 µl	0.07516 µl	± 0.50 µl	±	5.1E-02 µl
50.0252 µl	50 µl	-0.02522 µl	± 0.50 µl	±	7.9E-02 µl
100.0905 µl	100 µl	-0.09053 µl	± 0.50 µl	±	9.6E-02 µl