

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.298 s	3602 s	1.702 s	0.047%	± 0.01 %	± 5.6E-02 s
3599.862 s	3599 s	-0.862 s	-0.024%	± 0.01 %	± 5.6E-02 s
3599.325 s	3599 s	-0.325 s	-0.009%	± 0.01 %	± 5.6E-02 s