

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0033 g/cm³	0.997 g/cm³	-0.00629 g/cm³	± 0.01 g/cm³	±	8.9E-02 g/cm³