

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1,525.39 ml	1525.40 ml	0.01294 ml	± 0.1 ml	±	7.6E+01 µl