

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.0 °C	-1.062 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.038 %HR	40.0 %HR	9.962 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	40.0 %HR	-10.437 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	40.0 %HR	-30.028 %HR	± 5.0 %HR	± 1.9E+00 %HR