

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.038 %HR	34.5 %HR	4.462 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	50.7 %HR	0.263 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	69.2 %HR	-0.828 %HR	± 3.0 %HR	± 1.9E+00 %HR