

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.101 °C	0.0 °C	0.101 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.881 °C	15.1 °C	0.219 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.551 °C	29.8 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.496 %HR	37.1 %HR	2.604 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.546 %HR	49.6 %HR	1.054 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.235 %HR	68.3 %HR	-2.935 %HR	± 3.0 %HR	± 1.9E+00 %HR