

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.101 °C	-0.4 °C	-0.299 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.881 °C	14.7 °C	-0.181 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.551 °C	30.0 °C	0.449 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.496 %HR	37.2 %HR	2.704 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.546 %HR	50.7 %HR	2.154 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.235 %HR	68.9 %HR	-2.335 %HR	± 3.0 %HR	± 1.9E+00 %HR