

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.101 °C	-0.1 °C	0.001 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	14.881 °C	15.4 °C	0.519 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.551 °C	30.1 °C	0.549 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.496 %HR	32 %HR	-2.496 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.546 %HR	47 %HR	-1.546 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.235 %HR	74 %HR	2.765 %HR	± 5.0 %HR	± 1.9E+00 %HR