

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.101 °C	1.1 °C	1.201 °C	± 1.0 °C	± 4.0E-01 °C
15 °C	14.881 °C	16.1 °C	1.219 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.551 °C	30.7 °C	1.149 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.496 %HR	31 %HR	-3.496 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.546 %HR	44 %HR	-4.546 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.235 %HR	67 %HR	-4.235 %HR	± 5.0 %HR	± 1.9E+00 %HR