

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.2 μm	24.2 μm	1.0 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	51.0 μm	-0.1 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	124 μm	-1.1 μm	± 2.5 μm	±	6.1E-02 μm
478.0 μm	479 μm	1.0 μm	± 2.5 μm	±	9.3E-02 μm
952 μm	951 μm	-1.0 μm	± 2.5 μm	±	1.6E-01 μm