

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	24.8 μm	1.6 μm	± 2.5 μm	±	5.8E-02 μm
51.1 μm	51.6 μm	0.5 μm	± 2.5 μm	±	5.8E-02 μm
125.1 μm	122 μm	-3.1 μm	± 2.5 μm	±	6.1E-02 μm
478.0 μm	475 μm	-3.0 μm	± 2.5 μm	±	9.3E-02 μm
952 μm	961 μm	9.0 μm	± 2.5 μm	±	1.6E-01 μm