

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 99.98 RPM                | 100 RPM          | 0.02 RPM             | ± 1 RPM                     | ±                                  | 2.1E-01 RPM |
| 999.64 RPM               | 1000 RPM         | 0.36 RPM             | ± 1 RPM                     | ±                                  | 6.6E-01 RPM |
| 9999.26 RPM              | 10000 RPM        | 0.74 RPM             | ± 1 RPM                     | ±                                  | 2.0E+00 RPM |
| 13999.20 RPM             | 14000 RPM        | 0.80 RPM             | ± 1 RPM                     | ±                                  | 2.6E+00 RPM |