

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
7 Hz	0.400 g-pk	0.38 g-pk	-0.02 g-pk		±	0.0093 g-pk
10 Hz	0.800 g-pk	0.82 g-pk	0.02 g-pk		±	0.0163 g-pk
50 Hz	1.000 g-pk	1.02 g-pk	0.02 g-pk		±	0.0203 g-pk
100 Hz	1.000 g-pk	1.02 g-pk	0.02 g-pk		±	0.0207 g-pk
500 Hz	1.000 g-pk	1.02 g-pk	0.02 g-pk		±	0.0204 g-pk
1000 Hz	1.000 g-pk	0.97 g-pk	-0.03 g-pk		±	0.0209 g-pk