

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.01585 ml	50 ml	-0.01585 ml	± 0.04 ml	±	3.5E-01 µl