

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.182 s	3600 s	-0.182 s	-0.005%	± 0.01 %	±	5.6E-02 s
3600.702 s	3600 s	-0.702 s	-0.019%	± 0.01 %	±	5.6E-02 s
3600.512 s	3600 s	-0.512 s	-0.014%	± 0.01 %	±	5.6E-02 s