

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.401 °C	0.6 °C	0.199 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.215 °C	20.7 °C	0.485 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.126 °C	35.6 °C	0.474 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	30.6 %HR	0.389 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.672 %HR	55.0 %HR	-0.672 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.059 %HR	78.9 %HR	-1.159 %HR	± 3.0 %HR	± 1.9E+00 %HR