

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.401 °C	0.4 °C	-0.001 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.215 °C	19.8 °C	-0.415 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.126 °C	34.8 °C	-0.326 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	32.3 %HR	2.139 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.672 %HR	53.9 %HR	-1.772 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.059 %HR	79.4 %HR	-0.659 %HR	± 3.0 %HR	± 1.9E+00 %HR