

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.401 °C	0.3 °C	-0.101 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.215 °C	20.0 °C	-0.215 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.126 °C	34.7 °C	-0.426 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	31.0 %HR	0.839 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.672 %HR	55.8 %HR	0.128 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.059 %HR	78.0 %HR	-2.059 %HR	± 3.0 %HR	± 1.9E+00 %HR