

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	29.9 °C	0.058 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.842 %HR	31.2 %HR	2.358 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.435 %HR	49.4 %HR	-0.035 %HR	± 2.5 %HR	± 1.9E+00 %HR
75 %HR	75.468 %HR	73.9 %HR	-1.568 %HR	± 2.5 %HR	± 1.9E+00 %HR