

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	29.5 °C	-0.342 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.957 %HR	32.5 %HR	2.543 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	49.435 %HR	51.8 %HR	2.365 %HR	± 2.5 %HR	± 1.9E+00 %HR
75 %HR	75.468 %HR	76.8 %HR	1.332 %HR	± 2.5 %HR	± 1.9E+00 %HR