

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1558.4 RPM	1500 RPM	-58.4 RPM	± 5 RPM	±	1.2E+00 RPM
2072.6 RPM	2000 RPM	-72.6 RPM	± 10 RPM	±	1.9E+00 RPM