

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
759.84 RPM	600 RPM	-159.84 RPM	± 753 RPM	±	4.7E+00 RPM
280.58 RPM	300 RPM	19.42 RPM	± 276.8 RPM	±	8.5E-01 RPM
251.80 RPM	200 RPM	-51.80 RPM	± 251.7 RPM	±	4.7E-01 RPM
126.54 RPM	100 RPM	-26.54 RPM	± 126.4 RPM	±	1.6E-01 RPM