

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
194.58 RPM	180 RPM	-14.58 RPM	± 194 RPM	±	6.5E-01 RPM
1087 RPM	1000 RPM	-87 RPM	± 1090 RPM	±	3.7E+00 RPM
1609.80 RPM	1500 RPM	-109.80 RPM	± 1616 RPM	±	9.9E-01 RPM