

| Patrón <i>(Standard)</i> | UBP <i>(UIT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |             |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|-------------|
| 194.58 RPM               | 180 RPM          | -14.58 RPM           | ± 194 RPM                   | ±                                  | 6.5E-01 RPM |
| 1087.00 RPM              | 1000 RPM         | -87.00 RPM           | ± 1090 RPM                  | ±                                  | 3.7E+00 RPM |
| 1609.80 RPM              | 1500 RPM         | -109.80 RPM          | ± 1616 RPM                  | ±                                  | 9.9E-01 RPM |