

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
18 °C	18.130 °C	18.1 °C	-0.030 °C	± 1.6 °C	± 4.0E-01 °C
23 °C	23.143 °C	23.2 °C	0.057 °C	± 1.6 °C	± 4.0E-01 °C
28 °C	28.157 °C	28.3 °C	0.143 °C	± 1.6 °C	± 4.0E-01 °C
45 %HR	45.376 %HR	48.0 %HR	2.624 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.640 %HR	63.2 %HR	3.560 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.141 %HR	77.2 %HR	2.059 %HR	± 5.0 %HR	± 1.9E+00 %HR