

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.038 %HR	33.0 %HR	2.962 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	53.2 %HR	2.763 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	71.2 %HR	1.172 %HR	± 3.0 %HR	± 1.9E+00 %HR