

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.774 °C	1.9 °C	0.126 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.730 °C	3.8 °C	0.070 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.711 °C	7.8 °C	0.089 °C	± 0.5 °C	± 4.0E-01 °C