

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.035 °C	19.9 °C	-0.135 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	27.957 °C	27.7 °C	-0.257 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.2 °C	-0.291 °C	± 0.5 °C	± 4.0E-01 °C