

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100 RPM	100 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
150 RPM	150 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
210 RPM	210 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
250 RPM	250 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
300 RPM	300 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
350 RPM	350 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
420 RPM	420 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
490 RPM	490 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
560 RPM	560 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
630 RPM	630 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM
700 RPM	700 RPM	0 RPM	± 1 RPM	±	5.8E-01 RPM