

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99956	0.440	± 1	±	4.3E-01 mm						
0 a 2	1.99985	0.150	± 1.4	±	4.8E-01 mm						
0 a 3	2.99943	0.570	± 1.8	±	4.8E-01 mm						
0 a 4	3.99959	0.410	± 2.2	±	4.8E-01 mm						
0 a 5	4.99954	0.460	± 2.6	±	4.8E-01 mm						