

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	# 0 mm	0.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2.5 mm	2.50 mm	0 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	5.1 mm	5.10 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	7.7 mm	7.70 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10.3 mm	10.30 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	12.9 mm	12.89 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15.0 mm	15.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	17.6 mm	17.61 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20.2 mm	20.21 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	22.8 mm	22.8 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25.0 mm	25.0 mm	10 µm	± 30 µm	± 6.1E+00 µm