

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	# 0.0 mm	0.05 mm	50 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	2.5 mm	2.53 mm	30 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	5.1 mm	5.14 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	7.7 mm	7.75 mm	50 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	10.3 mm	10.34 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	12.9 mm	12.93 mm	30 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	15.0 mm	15.04 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	17.6 mm	17.64 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	20.2 mm	20.24 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	22.8 mm	22.8 mm	40 μm	± 50 μm	± 2.9E+01 μm
Profundidad / (Depth)	25.0 mm	25.0 mm	40 μm	± 50 μm	± 2.9E+01 μm